aidacare™
HEALTHCARE EQUIPMENT

200S STEEL STANDING LIFTER

USER MANUAL



Congratulations and thank you for choosing an Aidacare 200S Standing Lifter.

This product has been designed to offer maximum safety for the user and reduced manual handling for the carer.

Aidacare standing lifters are for Indoor use only.

IMPORTANT: Do not use the Aidacare 200S Standing Lifter without having first read this user manual

This product Features the UKCA mark, in compliance with Part II UK MDR 2002 (as amended) Class 1.

This product features the CE mark, in compliance with the Medical Devices Regulation 2002 (as amended) Class 1, and the Medical Device Regulation 2017/745

© Aidacare Limited 2026

No part of this user manual may be copied, distributed, reproduced, scanned, or stored on any electronic database, whether in whole or in part in any form or by any means.

Due care has been taken to ensure all the information contained in this user manual is correct at the time of printing. All measurements, pictures, colours, and weight capacities are to be used as a guide only. We reserve the right to modify the design or appearance of any product displayed in this user manual without prior notice.

TABLE OF CONTENTS

1.	Symbols Used In This Manual	4
2.	Symbols Used On The Product	5
3.	Overview	5
4.	Safety Precautions & Warnings	6
5.	Intended Use	9
6.	Parts Glossary	11
7.	Assembly	12
8.	Preliminary Checks	14
9.	Standing Lifter Operation	14
10.	Standing Safe Operating Instructions	19
11.	Sling Compatibility	29
12.	Care & Maintenance	30
13.	Compliance	31
14.	Warranty & Service	32
15.	Troubleshooting	33
16.	Specifications	35
17.	Recycling Instruction	36
18.	Manufacturer, UK Representative & Importer Details	37
19.	Service Log Book	38

1. SYMBOLS USED IN THIS MANUAL

The symbols below are used throughout this user manual and on the product to identify warnings and important information. It is very important for you to read them and understand them completely.



WARNING! Indicates a potentially hazardous condition/situation. Failure to follow designated procedures can cause either personal injury, component damage or malfunction. On the product, this icon is represented as a black symbol on a white triangle with a black border.




ALWAYS! These actions should be performed as specified. Failure to perform mandatory actions can cause personal injury and/or equipment damage. On the product, this icon is represented as a white infinity symbol on a black dot with a white border.



DO NOT! These actions are prohibited. These actions should not be performed at any time or in any circumstances. Performing a prohibited action can cause personal injury and/or equipment damage. On the product, this icon is represented as a white symbol with a black circle and black slash.

2. SYMBOLS USED ON THE PRODUCT



Please note: Sticker is sample only and not to scale

# Product Code	UK CA UKCA Mark
SN Serial Number	SWL 200KG Safe Working Load
Manufacturer	Read User Manual
Importer	Caution
CN Country of Manufacture	MD Medical Device
Class 2 Electrical Equipment	EC REP Authorised Representative in EU
Type B as per IEC60601-1	Includes Electronics components and should be mixed with General household waste
CE CE Mark	

3. OVERVIEW

The **Aidacare 200S Standing Lifter** is an ergonomically designed mobile standing lifter with a generous safe working load (SWL) of 200kg / 31.5st. This mobile standing lifter is ideally suited as a familiar and reliable solution for institutional and aged care environments.

The 200S Standing Lifter allows patients to be transferred from sitting to standing and vice versa with comfort and safety.

The proven bull horn shape yoke offers versatility in patient grip points and multiple sling hooks to meet various user and carer needs. The height and width adjustable knee pads offer lower leg comfort and security during the lift.

Our range of patient handling solutions feature industry leading LINAK Jumbo Care electronics and have been designed and tested to comply with ISO 10535:2021.

4. SAFETY PRECAUTIONS & WARNINGS

SAFETY CONSIDERATIONS



ALWAYS ensure that attendants have read this guide before operating this lifter.



ALWAYS use slings that are assessed and considered suitable by a trained healthcare professional.



ALWAYS ensure the area around the lifter and patient is clear of obstructions.



ALWAYS minimise the distance over which the patient is being transferred by positioning support surfaces as close as possible.



DO NOT operate the lifter without training in the safe operation of lifters and slings. **IF IN DOUBT DO NOT USE**.



DO NOT use the lifter where any doubt exists regarding the safety of the standing lifter including but not limited to wear or damage with particular importance to be placed on the yoke and lifter sling attachment points.



DO NOT use the lifter with slings showing any signs of wear or damage, including fraying.



DO NOT exceed the 200kg / 31.5st safe working load of the Aidacare 200S Lifter.



DO NOT use the lifter as a means of transportation over distances in excess of 3 metres. It is not designed as a transportation device.



DO NOT use the lifter if any damage, fault, or abnormal operation is identified. Refer to ‘Withdrawal from Service’ criteria in the Care & Maintenance section.

INCIDENT REPORTING



WARNING! Any incident, malfunction, or near-miss involving the lifter must be reported immediately to the equipment owner, facility manager or authorised Aidacare representative.

Incidents requiring reporting include, but are not limited to:

- Patient falls or slips during lifting or transfer
- Sling detachment or incorrect sling engagement
- Unexpected descent or uncontrolled boom movement
- Failure of pre-use safety checks
- Activation of emergency lowering functions during use
- Electrical faults or charging failures
- Overload events or suspected overload
- Unusual noises, instability, or abnormal equipment operation
- Any event resulting in injury or potential injury to the patient or carer

Remove the lifter from service until the equipment has been inspected and declared safe for continued use.

BATTERY & CHARGER SAFETY WARNINGS



WARNING! Immediately stop using the lifter, battery, or charger and disconnect from mains power if any of the following conditions are observed:

- Battery overheating or excessive heat
- Smoke, sparks, or unusual odours
- Swelling, cracking, or deformation of the battery casing
- Liquid leakage or signs of fluid ingress
- Damaged charger housing, plugs, or exposed wiring
- Abnormal charging behaviour or failure to charge correctly
- Evidence of burning, melting, or discolouration
- **DO NOT** attempt to open, repair, modify, or continue using damaged batteries or chargers.
- Only use batteries and chargers approved by Aidacare for use with this lifter.
- Damaged electrical components must be replaced by an authorised service technician prior to further use.

5. INTENDED USE

The 200S Standing Lifter has been designed to stand and transfer human patients with the assistance of one carer or attendant. This product is suitable for indoor areas only and it should **NEVER** be used outdoors.

All Aidacare standing lifters should **ONLY** be used with Aidacare approved patient slings, which have been chosen and assessed by a competent trained health professional. For a complete list of all compliant lifting slings please contact Aidacare.

WARNING

IMPORTANT – only approved slings and accessories are compatible with this lifter. If you are unsure please contact Aidacare.

The correct and appropriate use of the 200S Standing Lifter reduces the risks associated with manual patient handling. The lifting and transferring of patients should always be assessed and planned by a trained health professional.

This lifter is designed for indoor use **ONLY** and should **NOT** be used outdoor under any circumstances.

WARNING

WARNING: DO NOT manually lift the boom of lifters.



This is the direction of travel for which this standing lifter has been designed.

How to Use Correctly:

The boom of the lifter is designed to be lifted only by an electric actuator.

Aidacare Lifters and Standing Lifters do not pull downwards during lowering. This safety spline feature is designed to avoid entrapment and crush injuries.

This safety feature means that it is physically possible to raise the lifter boom by hand on some models with LA31 actuators fitted.

DO NOT raise the lifter boom by hand or any other means except under direct drive from the electric actuator fitted to the lifter.


Raising the boom can cause it to become stuck in an elevated position. Under load this can result in the patient to free fall and risks catastrophic injury to the patient and/or carer.

ALWAYS use the handset or control unit buttons to raise and lower the boom of the lifter. This will ensure correct and safe functioning at all times.



6. PARTS GLOSSARY






1	Boom	9	Push Handle
2	Sling Bar / Yoke	10	Adjustable Knee Pads
3	Lift Actuator	11	Rear Braking Castor
4	Mast	12	Foot Platform
5	Battery	13	Front Castors
6	Cover for Leg Actuator	14	Base Fastening Lever
7	Leg Spread Actuator	15	Emergency Lowering Function
8	Handset	16	Emergency Stop Button

7. ASSEMBLY



1. Unpack

Remove all parts from carton.





2. Attach the mast

Insert the mast into the base, ensuring the holes align and secure with base fastening lever bolt.






3. Apply thread locker / loctite


a) Locate actuator mounting hardware.

b) Install actuator hardware. A few drops of medium strength thread locker / loctite must be applied to the bolt threads.



4. Mount handle

Mount push handle onto mast in position shown and fasten screws with Allen key.



5. Attach the knee pads

Position the knee pads onto the mount point located toward the base of the standing lifter. Attach with four fixing bolts (supplied)





6. Connecting the actuator

a) Connect grey actuator lead into port (1) located on base of the control unit.

b) Connect the other end into the female lead from the actuator.



7. Connecting the handset

Connect grey handset lead into port located on base of control unit.



8. Connecting the Leg Spread Actuator

Connect black actuator lead port (2) on base of control unit.

8. PRELIMINARY CHECKS

After charging the battery these checks should be completed prior to initially using the standing lifter:

1. Check handset functionality is correct.
2. Check to see that all indicator lights are functioning.
3. Check that the emergency stop button causes the standing lifter to stop.
4. Check emergency lowering function.
5. Check that all leads are inserted, fastened and clear of the rear castors.

9. STANDING LIFTER OPERATION

Handset

Features:

1. Raise / Lower Function Buttons

   Press to move the standing lifter up and down. Release to stop movement. Always ensure the environment is free of obstacles that may impede movement.

2. Leg Spread Function Buttons

   Press to open and close the standing lifter's legs. Release to stop movement. This allows you to fit the standing lifter around furniture, such as chairs. Always ensure the environment is free of obstacles that may impede movement.

3. Battery Charge Indicators

   Battery charge indicators will flash when battery reaches maximum recommended discharge of 18 volts, while the standing lifter is under load.

4. Service Indicator

   This will illuminate to show that a service is required. This is usually after 12 months or 8,000 lift cycles – whichever comes first. You do not have to stop using your standing lifter when this illuminates, simply contact your Aidacare dealer to arrange a service.



- Press and hold the raise / lower or leg spread function buttons to activate. Release to deactivate. Beware of obstacles at all times to ensure boom and legs can move freely.
- Battery charge indictors will flash when battery reaches maximum recommended discharge of 18 volts, while the standing lifter is under load.
- Service indicator will illuminate when standing lifter has reached recommended service interval, either 12 months since last service or 8000 lift cycles. At this time the standing lifter needs to be serviced for safety and to maintain applicable warranty.

A full load test should be completed at least every 12 months by an Aidacare service technician.



Emergency Lowering


To activate emergency lowering lift up red tab located on top of actuator housing.


This allows the patient to be fully lowered in the event of a fault in the standing lifter or discharged batteries.

Kneepad Adjustment

Adjust kneepad height by loosening hand wheel bolt on upright post. Set to desired height and fasten back in place.

Adjust knee pad width by releasing spring pins on each side, allowing the pin to auto-locate into the preset positions. Set to desired width and fasten back in place.





Charging the battery

The 200S Standing Lifter is supplied with a LINAK external wall mount charger and is compatible with standard Australian power outlets.

The following steps should be completed to recharge the battery pack correctly:



a) Remove the battery pack from the lifter via the top mounted release lever.



b) Connect the battery pack to the charger.

c) Charger will automatically activate the charging sequence and a light will illuminate to show charging has commenced.

CHARGING INDICATOR DISPLAY

GREEN LIGHT indicates mains power is connected to 240V power supply.

ORANGE LIGHT indicates battery is charging, and will turn off automatically when battery is fully charged.



d) When fully charged, place the battery back onto the standing lifter, ensuring that you hear a click. This indicates the battery is sitting correctly in the standing lifter.

CHARGING DO'S AND DON'TS


ALWAYS

- Charge the battery whenever possible, to maximise battery life. Allowing batteries to fully discharge can reduce the useful life of a battery.
- Always inspect all cables particularly the mains power cable on the charger for any damage; replace if any signs of damage or wear are visible.
- Always stow the handset, charger and mains power cable when transporting the lifter to prevent loss and or damage.
- Always clean the actuators, control box, charger, battery and handset at regular intervals to remove dust/dirt.

NEVER

- Never allow the battery to fully discharge before recharging. Allowing batteries to fully discharge can reduce the useful life of a battery.

Castors



To **engage the castor brake** place your foot on the brake pedal and apply sufficient downward force to move it into the lock position.



To **disengage the castor brake** simply place the point of your toe underneath the brake pedal and lift-up.

Emergency Stop

To engage the emergency stop press the red button located on the control unit until it sits flush and remains in position.





To disengage the emergency stop function twist the red button clockwise until it pops out




10. STANDING LIFTER (SIT-TO-STAND HOISTS) SAFE OPERATION INSTRUCTIONS

All lifter operators should be trained in safe manual handling and be familiar with the correct use of the lifter and sling before assisting with a transfer. Each user should be assessed by a suitably qualified health professional to ensure the appropriate sling and transfer method are selected, and the lift process should be clearly explained with the user's consent obtained to support comfort, confidence, and cooperation. It is strongly recommended that transfers should be performed with the assistance of two carers to enhance safety.


Standing lifters are designed to assist a user who can actively participate and bear some weight through their lower limbs to transfer between seated and standing positions (e.g. chair to toilet, bed to chair).

WARNINGS




WARNING

Lifter legs
Opening the legs prior to lifting or lowering increases the base of support and maximises stability.



WARNING

Brakes
Lifter brakes should be **OFF** when the user is being lifted and lowered by the hoist.



WARNING

User's hand placement
Ensure the operators and users fingers and hands are away from the loops, hooks, pins, and clips as these are all potential pinch points.

WARNING

Safe Working Load (SWL)
Never exceed the SWL of the lifter or sling.

WARNING

Supervision
Never leave a user unattended while supported by the lifter.

PRE-USE CHECKS

Risk assessment: Confirm that the task, environment, and user are suitable for the transfer.

Inspect lifter: No loose, missing, or damaged parts. Check that the castors move freely and that the brakes engage and release. Check that the battery is adequately charged, the emergency stop is working, and the handset is functioning. Confirm safe working load of the lifter.

Inspect sling: Check for damage. Review information labeling to confirm safe working load, type, size, and material. Ensure the sling is compatible with the lifter.


STEP 1 Positioning the User

The user should start from a seated position. If they are able to do so, ask the user to lean forward or assist them to do so using appropriate manual handling techniques.




STEP 2 Sling Application

Position the sling down behind the user's back until the underarm pads are just below the armpits.




Pull the waist belt and strap loops or clips through under the user's arms.




Ensure that the sling is centrally aligned and strap length is the same on both sides.


Fasten the velcro first, firm around the user's waist, and then further secure it with the buckle. It should be firm but not tight around the user's waist.



NOTE: For comprehensive instructions on sling application, refer to the user manual for the specific sling model being used.

STEP 3 Bring the lifter into position






Open the lifter legs to maximise stability, if applicable.

Lower the boom and move the lifter in front of the user so that it is at chest level, or low enough so that the sling loops or clips can easily be attached to the yoke hooks or clip sling bar pins.




WARNING! Take care to ensure the sling bar does not come into contact with the user's face or body causing incidental harm.


NOTE: It may be helpful to use the functions of the adjustable bed, recliner, or similar to make attaching the sling easier. However, care must be taken to ensure equipment is not adjusted when the standing lifter yoke/clip boom frame is close to the user as this could result in accidental contact and potential injury.




Correct position for yoke or clip boom frame at chest height




Incorrect position of yoke or clip boom




Yoke or clip boom too close or in contact with the user





Ask the user to lift their feet onto the footplate or assist them to do so.





Adjust the knee pads. The top of the knee pad should rest just below the user's kneecap.






STEP 4 Attaching the Sling


Attach the sling to the yoke/clip boom. Firmly pull on the sling to ensure correct engagement of the loops or clips.

NOTE: For slings with multiple connection points, position should be determined based on users size/height. Once connected, the sling straps should remain taut and allow the user to lean back into the sling slightly.


YOKE AND LOOP SLING




Loop hooks sitting taut in the yoke with the slings closed




Loop sitting on top of the hook






Only one loop connected




Twisted loops




Loop length (loop colours) do not match


CLIP BOOM AND CLIP SLING




Clip sling securely attached onto pin



Clip sitting outside key hole mechanism and not locked into pin, OR using the loop part of the clip sling



Same clips attached on both sides




First clip connected on one side and second clip connected on the other




Only one clip connected

WARNING


WARNING! Before commencing the lift, ensure the brakes are **OFF**.






STEP 5 Checking the Sling


Ask the user to hold onto the handles. Slightly raise the boom until the sling straps are taut but before the user's bottom leaves the support surface underneath.




Then STOP and CHECK
straps before continuing.




YOKE AND LOOP SLING




Straps are taut and loops are secure in hooks with clips closed. Ensure there are no twists in the loop straps




Sling loop colours do not match on both sides




Loop attached to one hook only




Clip missing or bent



Loop loose in hook




Loop on top of hook




Loop doubled over, strap not taut


CLIP BOOM AND CLIP SLING




Both clips are securely attached to the clip boom pins and locked into the key hole position




Only one clip attached




Clip cracked, broken or missing



Same clips attached on both sides



First clip connected on one side and second clip connected on the other




Clip over pin but not properly engaged


STEP 6 Carrying out the transfer

If the sling's loops or clips remain secure, the lift can be continued.


Encourage active effort, prompt the user to push through their legs as able.

NOTE: Always position the two support surfaces that the lift is taking place between as close together as possible. Remember this is a transfer tool, not a transport tool.






Two pieces of equipment side by side



Two pieces of equipment more than 3m apart

STEP 7 Lowering onto destination equipment

Once the user is upright, using appropriate manual handling techniques, manoeuvre the lifter to the piece of equipment being transferred to. Once positioned correctly, lower the user until they are fully supported on the destination surface.



WARNING

WARNING! The brakes of the equipment being transferred to should be **ON**. e.g. bed/wheelchair/commode.


STEP 8 Sling and lifter removal

Lower the boom so that the sling straps are slack. Remove the sling from the yoke and move the lifter out of the way before removing the sling from the user.

11. SLING COMPATIBILITY

The Deluxe Padded Standing Sling has been developed to work seamlessly with the Aidacare range of standing lifters.

This standing lifter may also be suitable for use with other brands of slings. To confirm if a sling is compatible with any of the Aidacare range of patient lifting systems contact Aidacare.



Deluxe Padded Standing Sling

12. CARE & MAINTENANCE

To prolong the working life of the standing lifter and for trouble free operation it is recommended that periodic inspections should be performed routinely by a person who is suitably qualified and well acquainted with the design, use and care of patient standing lifters.

If inspection reveals that the safety of the standing lifter is jeopardized in any way the standing lifter should be **IMMEDIATELY WITHDRAWN FROM SERVICE**, tagged as such and the owner notified.

COMMON CHECKS, MAINTENANCE AND CLEANING

- Check castors are fastened securely to the frame
- Castors should run and swivel smoothly – if hair or other fibres have built up within the castor this should be removed
- Check that the brakes on the rear castors function normally
- Clean surfaces with a pH neutral detergent – chlorine and phenol based detergents should not be used as these may damage standing lifter components
- All moving parts and pivot points should be checked for fatigue and lubricated if necessary
- All electrical components should be checked for normal function including battery and battery charger
- The actuator shaft should only be wiped with a cloth. **DO NOT** lubricate. Should you have any questions, contact Aidacare.

AS A MINIMUM, A FULL SERVICE AND LOAD TEST SHOULD BE COMPLETED EVERY 12 MONTHS TO MAINTAIN WARRANTY.

ELECTRICAL TEST & TAG – WHERE STANDING LIFTERS ARE IN USE IN A COMMERCIAL ENVIRONMENT THEN THE TESTING REGIME APPLIED SHOULD ALIGN WITH THE RISK ASSESSMENT MADE BY THE OWNER FOR SAFETY AND OTHER COMPLIANCE REQUIREMENTS.

TO MAINTAIN GLOBAL STANDARDS COMPLIANCE UNDER AS ISO 10535-2021 A FULL SWL TEST MUST BE COMPLETED EVERY 12 MONTHS.

WITHDRAWAL FROM SERVICE



WARNING! Immediately remove the lifter from service and isolate the equipment if any of the following conditions are identified:

- Cracked, bent, deformed, or damaged structural components
- Loose or missing fasteners, pins, hooks, or retaining devices
- Damaged, worn or failed yoke, spreader bar, or sling attachment points
- Failed, damaged, or ineffective brakes or castors
- Electrical faults, exposed wiring, or damaged connectors
- Failure of emergency stop or emergency lowering functions
- Unexpected boom movement or unintended descent
- Unusual noises, vibration, or abnormal operation
- Unreadable, missing or damaged safety labels or load ratings
- Evidence or suspicion of overload or misuse
- Damage caused by impact, corrosion, liquid ingress or improper storage



DO NOT use the lifter until it has been inspected and approved for use by an authorised Aidacare service technician.

13. COMPLIANCE

This standing lifter is manufactured in ISO compliant production facilities in PR China to strict quality control standards.

Aidacare standing lifters are manufactured and independently tested to the market relevant compliance standards.

Labels on the standing lifter address compliance requirements and users should familiarise themselves with these prior to use.

14. WARRANTY & SERVICE

STANDER AND ELECTRONICS

Aidacare warrants its products to be free from defects in materials and workmanship under normal use and service and will within the period stated below, from the date of purchase, repair or replace without cost to the original customer, any part assembly or portion thereof which shall be returned to Aidacare and from OUR inspection shows to be defective.

1. Standing Lifter Frame – 5 years
2. LINAK Actuator and LINAK Controller – 3 years (excluding Battery Pack)
3. LINAK Battery Pack & Handset – 12 month manufacturer's warranty subject to stringent fair wear and tear clauses (please contact AIDACARE for a copy of this information)

WARRANTY IS CONTINGENT UPON ANNUAL SERVICING BEING COMPLETED FOR THE LENGTH OF THE WARRANTY PERIOD.

Aidacare cannot be held responsible for any personal injury, damage to the standing lifter or damage to property as a result of the improper or unsafe use of the product.

No warranty claim shall apply where the product or any part thereof has been modified, varied, altered or damaged either accidentally or through improper or negligent use.

No warranty claim shall apply where the standing lifter is repaired or serviced by parties other than those accredited and or approved by Aidacare.

Warranty does not extend to items or components which may require replacement due to normal wear and tear (e.g. castors, plastic molded components, foam padding (where fitted) and painted surfaces) for which Aidacare nor its distributors can be held responsible.

BATTERIES

Batteries carry a limited warranty from the original manufacturer which is subject to a stringent wear and tear clause. Battery faults due to defect in original manufacture typically become apparent within the first two months of use.

Any gradual deterioration in performance after this period is typically associated with fair wear and tear, misuse and accidental damage, and as such is not covered by the LINAK manufacturer's warranty.

ALL WARRANTIES MADE ARE BACK TO BASE WARRANTIES MEANING THAT ANY PARTS OR WHOLE STANDERS REQUIRING REPAIR NEED TO BE RETURNED TO THE NEAREST AIDACARE SERVICE CENTRE AT THE CUSTOMER'S EXPENSE – PLEASE CONTACT AIDACARE FOR A CURRENT LIST OF SERVICE CENTRE LOCATIONS.

SPARE PARTS

SPARE PARTS LIST AVAILABLE UPON REQUEST FROM AIDACARE

CUSTOMISATION AND MODIFICATION

All lifter customisation, modifications and or alterations made to this product are considered as "custom" changes. These may cause the product to no longer meet safety standards and should be considered prior to commencement.

Any customisation or modification of the lifter is not warranted by Aidacare and will void the warranty.

15. TROUBLESHOOTING



WARNING: TROUBLESHOOTING SHOULD ONLY BE CARRIED OUT BY AIDACARE AUTHORISED SERVICE TECHNICIANS WHO ARE TRAINED IN SERVICING THE ELECTRICAL AND MECHANICAL FEATURES OF THE PATIENT STANDER.

The following is a guideline only to general troubleshooting.

Before carrying out any troubleshooting:


- CHECK TO ENSURE the RED emergency stop button is NOT pushed in.
- CHECK TO ENSURE maintenance procedures have been followed.
- CHECK TO ENSURE all lead connections are secure.
- CHECK TO ENSURE the battery is fully charged.


PROBLEM	POTENTIAL FAULT	ACTION
• Warning alarm can be heard in control unit	• Battery is discharged • Battery may be faulty	• Recharge battery • Replace battery
• Power LED on charger does not illuminate	• Not connected to power • Fuse blown • Faulty Charger	• Check plugged in to power point and switched on • Replace fuse • Repair or replace charger
• Actuator not running/ no click heard from control unit when buttons pressed	• Faulty control unit • Faulty handset	• Repair or replace control unit
• Actuator not running but control unit clocks when buttons pressed	• Actuator plug not inserted fully into control unit • Faulty actuator • Faulty control unit	• Check plugs are completely inserted into control unit ports • Repair or replace actuator and or control unit
• Actuator not lifting properly or running slowly	• Discharged or faulty battery • Weight in excess of safe working load (SWL) • Faulty actuator	• Recharge or replace battery • Repair or replace actuator
• No clicking in control box when buttons pressed	• Emergency stop button pressed • Battery completely discharged • Battery fuse faulty • Faulty control unit	• Release emergency stop button • Recharge battery • Repair or replace control unit
• Actuator only works in one direction	• Faulty handset • Faulty control unit • Emergency lower release activated	• Repair or replace handset • Repair or replace control unit • Check emergency actuator release is reset

16. SPECIFICATIONS

Product Code	LSS390650	LSS390650 – UK
Safe Working Load (SWL)	200kg	31.5st
Base Length	1277mm	50.3"
External Base Width	665 – 1203mm	26" – 47.5"
Internal Base Width	561 – 1103mm	22" – 43.5"
Base Width Adjustment	Electric	
Base Height / Under Bed Clearance	110mm	4"
Minimum Height of Spreader Bar	857mm	34"
Maximum Height of Spreader Bar	1777mm	70"
Standing Lifter Weight	43.5kg	95.9lb
Gross Weight	62.4kg	137.6lb
Warranty	Frame: 5 Years Electronics: 3 Years Battery Pack: 1 Year	
Under Base Height	50mm	2"
Maximum Overall Height	1832mm	72"
Front Castor Diameter	75mm	3"
Rear Locking Castor Diameter	100mm	4"
Turning Diameter	1361mm	53.5"
Weight Of Heaviest Part (Base Assembly)	19.6kg	43.3lb
Handle Height	1045mm	41"
Sound Level	<40Db@ 1m	
Electrical Approval	EN 60601-1-2:2007; EN 61000-3-2:2006; EN 6100-3-3:2008	

Box Contains: Base with 4 castor wheels, mast and boom with yoke, LINAK handset, LINAK battery charger, LINAK battery, LINAK controller, push handle. Screws, bolts, nuts, washers, plastic caps and tool kit.





Max. Load and Net Weight in kg, the rest of the measurements are in cm (unmarked tolerance +/ – 1cm)															
Item No.	Max. Load	A		B	B'	C	D	F	F'	E	L		W	N.W	Turning Diameter
		Max.	Min								Max.	Min.			
7203	200	183.2	121.2	127.7	79	66.5-120.3 (tolerance +/_ 2cm)	56.1-110.3 (tolerance +/_ 2cm)	11	5	104.5	177.7	85.7	53	43.5	136.1

17. RECYCLING INSTRUCTION

This product is fitted with electronic components and should not be mixed with general household waste. Please follow the diagram below to identify the materials on the product. Old batteries and electronics are to be taken to your nearest recycling centre for recycling or contact your retailer.



Lead Battery (Pb)

Waste of Electrical and Electronic Equipment (WEEE)

Metals

18. MANUFACTURER, UK REPRESENTATIVE & IMPORTER DETAILS

If you have any questions, you can contact:




Manufacturer:

Aidacare Pty Ltd

ABN 40 134 398 833

Building 3A, 1 Moorebank Avenue, Moorebank NSW 2170
Australia

1300 133 120 | product@aidacare.com.au



EU Representative:

European Healthcare & Device Solutions (Ireland) Ltd

Stratton House, Bishopstown Road, Cork T12 Y9TC Ireland

info@europeandevicesolutions.eu



Importer:

Aidacare Ltd

Registration 13785408

Arcadia House, Maritime Walk, Ocean Village,

Southampton SO14 3TL United Kingdom

01622 541235 | info@aidacare.co.uk

19. SERVICE LOG BOOK

Year 1 Date: _____

Service Type: _____

Condition Report: _____

_____

Inspected by: _____ Signature: _____

Action taken: _____ Date: _____

Year 2 Date: _____

Service Type: _____

Condition Report: _____

_____

Inspected by: _____ Signature: _____

Action taken: _____ Date: _____

Year 3 Date: _____

Service Type: _____

Condition Report: _____

_____

Inspected by: _____ Signature: _____

Action taken: _____ Date: _____

Year 4 Date: _____

Service Type: _____

Condition Report: _____

_____

Inspected by: _____ Signature: _____

Action taken: _____ Date: _____

Year 5 Date: _____

Service Type: _____

Condition Report: _____

_____

Inspected by: _____ Signature: _____

Action taken: _____ Date: _____

AIDACARE PTY LTD

Building 3A, 1 Moorebank Avenue,
Moorebank NSW 2170
Australia

1300 133 120 | aidacare.com.au

AIDACARE LIMITED

Arcadia House, Maritime Walk,
Ocean Village, Southampton SO14 3TL
United Kingdom

01622 541235 | aidacare.co.uk

MADE IN PRC

290526

aidacare™
HEALTHCARE EQUIPMENT